How do I PAT test power tool chargers and mobile phone chargers?

Class 1 chargers should have continuity to the DC plug, so clip your earth lead there. Class 2, again clip your earth lead there as this may well be the only exposed metal part. Should the charger contain a transformer, then a thorough visual inspection will suffice.

If you require more help, please contact us at https://www.seaward.com/de/enquiry/.

15-18 Bracken Hill, South West Industrial Estate, Peterlee, County Durham, SR8 2SW, United Kingdom
t: +44 (0) 191 586 3511 | f: +44 (0) 191 586 0227 | e: sales@seaward.com
seaward.com