



## **When testing a long cord with a PrimeTest 250, the test fails. Which is the best way to test this type of product?**

When testing a cord using the Class I test on the PrimeTest 250, the test will indicate a failed result. The PrimeTest 250 has a fixed earth continuity pass limit of 0.2 ohm. As it has not got a user defined setting and uses a calculation of  $(0.1 + R)\square$  it would automatically fail any value over 0.2 ohm's, and then as continuity is a failure the balance of the test sequence will be considered a failure.

Seaward's 5th Edition compliant PrimeTest 250+ now includes variable pass/fail limits for earth continuity and a long lead test function, allowing the competent person to calculate the added resistance based on cross-sectional area of the cable and length and override the failure based on the new calculation.

If you require more help, please contact us at <https://www.seaward.com/sitemap/enquiry/>.